
Midmark M11 - v1.0.5
Total Cycles: 3225

Name: RITTER2
Sterilizer ID: V2313228
12/15/2023 03:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.6 F 0.0
2:00 158.2 F 2.0
4:00 181.8 F 6.0
6:00 203.0 F 10.4
8:00 217.4 F 6.0
10:00 225.8 F 8.8
12:00 237.2 F 12.3
14:00 240.5 F 11.4
16:00 239.7 F 10.4
18:00 253.1 F 17.0
20:00 264.4 F 23.8
22:00 271.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.6 F 29.3
2:00 271.3 F 28.5
3:00 272.2 F 29.5
4:00 272.3 F 29.2

Min 270.9 F 28.4
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 220.8 F 0.0
10:00 192.4 F 0.0
15:00 188.2 F 0.0
20:00 183.6 F 0.0
25:00 180.6 F 0.0
30:00 178.7 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 22:01
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:11
CYCLE COMPLETE

12/15/2023 04:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____