
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007818
9:31 03/09/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:06
STERILIZING 3:06
136.0 C 220kPa 3:06
136.4 C 221kPa 3:36
136.4 C 223kPa 4:06
136.3 C 223kPa 4:36
136.4 C 222kPa 5:06
136.1 C 221kPa 5:36
136.4 C 220kPa 6:06
136.1 C 221kPa 6:36
136.1 C 221kPa 6:36
VENTING 6:36
AIR DRYING 7:29
CYCLE COMPLETE 7:42

Digital Signature #
508D3CB0D4E25A7F
