
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001457
8:21 06/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:42
STERILIZING 4:48
136.4 C 223kPa 4:48
136.4 C 222kPa 5:18
136.1 C 223kPa 5:48
136.4 C 223kPa 6:18
136.4 C 224kPa 6:48
136.3 C 223kPa 7:18
136.4 C 221kPa 7:48
136.3 C 223kPa 8:18
136.3 C 223kPa 8:18
VENTING 8:18
AIR DRYING 9:12
CYCLE COMPLETE 9:14

Digital Signature #
28BE783DA046B0DB
