
Midmark M11 - v1.0.0
Total Cycles: 1861

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2022 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 217.0 F 0.0
2:00 201.4 F 3.7
4:00 201.3 F 8.8
6:00 221.4 F 7.5
8:00 231.4 F 10.1
10:00 234.3 F 7.8
12:00 245.0 F 12.4
14:00 245.1 F 12.7
16:00 258.2 F 19.8
18:00 269.9 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 271.6 F 28.2
3:00 272.6 F 29.2
4:00 272.1 F 28.6

Min 271.4 F 28.1
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 228.6 F 0.0
10:00 216.1 F 0.0
15:00 228.1 F 0.0
20:00 234.2 F 0.0
25:00 229.1 F 0.0
30:00 230.0 F 0.0

Drying Complete

FILLING: 1:34
HEATING: 18:57
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:38
CYCLE COMPLETE

04/25/2022 10:55 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____