
Midmark M11 - v1.0.5
Total Cycles: 2326

Name: RITTER2
Sterilizer ID: V2313228
05/02/2023 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.3 F 0.0
2:00 114.7 F 0.5
4:00 137.0 F 3.0
6:00 168.7 F 6.8
8:00 192.9 F 11.5
10:00 211.9 F 5.2
12:00 225.0 F 9.8
14:00 231.8 F 9.1
16:00 239.8 F 12.1
18:00 238.9 F 10.7
20:00 252.3 F 17.3
22:00 263.8 F 24.2
24:00 271.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.4 F 29.9
2:00 271.5 F 29.4
3:00 272.6 F 30.0
4:00 271.8 F 29.3

Min 271.5 F 29.2
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 195.9 F 0.0
10:00 208.9 F 0.0
15:00 220.3 F 0.0
20:00 220.6 F 0.0
25:00 221.6 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:01
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:14
CYCLE COMPLETE

05/02/2023 02:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____