

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4753

Name: RITTER 1  
Sterilizer ID: V2178198  
07/22/2024 10:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.8 F 0.0  
2:00 156.5 F 2.3  
4:00 166.3 F 5.7  
6:00 186.7 F 9.2  
8:00 201.7 F 12.9  
10:00 213.1 F 4.5  
12:00 223.7 F 8.2  
14:00 233.9 F 12.2  
16:00 233.0 F 8.3  
18:00 244.4 F 13.4  
20:00 242.5 F 11.6  
22:00 253.8 F 17.3  
24:00 263.7 F 23.1  
26:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.8 F 29.1  
2:00 271.9 F 28.4  
3:00 272.1 F 28.9  
4:00 272.6 F 29.0

Min 271.5 F 28.3  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.3 F 0.0  
5:00 192.7 F 0.0  
10:00 196.3 F 0.0  
15:00 194.1 F 0.0  
20:00 199.9 F 0.0  
25:00 202.0 F 0.0  
30:00 196.6 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 26:21  
STERILIZING: 4:00  
VENTING: 3:58  
DRYING: 30:00  
TOTAL CYCLE: 01:05:18  
CYCLE COMPLETE

07/22/2024 11:15 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_