
Midmark M11 - v1.0.0
Total Cycles: 4552

Name: RITTER 1
Sterilizer ID: V2178198
05/14/2024 01:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.9 F 0.0
2:00 110.6 F 0.4
4:00 129.8 F 2.6
6:00 154.2 F 5.3
8:00 180.9 F 8.7
10:00 205.1 F 12.5
12:00 212.5 F 5.2
14:00 221.7 F 9.0
16:00 232.8 F 13.1
18:00 235.6 F 8.9
20:00 242.9 F 11.7
22:00 243.4 F 11.9
24:00 254.5 F 17.6
26:00 264.6 F 23.7
28:00 271.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 28.9
2:00 272.5 F 29.3
3:00 271.2 F 28.6
4:00 272.8 F 29.1

Min 271.2 F 28.6
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 222.3 F 0.0
10:00 220.4 F 0.0
15:00 237.1 F 0.0
20:00 221.9 F 0.0
25:00 200.8 F 0.0
30:00 194.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 28:39
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:07:21
CYCLE COMPLETE

05/14/2024 02:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____