

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 005916  
9:17 06/07/2023

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 2:44  
STERILIZING 3:42  
136.4 C 223kPa 3:42  
136.4 C 223kPa 4:12  
136.4 C 225kPa 4:42  
136.1 C 222kPa 5:12  
136.4 C 218kPa 5:42  
136.1 C 223kPa 6:12  
136.3 C 222kPa 6:42  
136.2 C 223kPa 7:12  
136.2 C 223kPa 7:12  
VENTING 7:12  
AIR DRYING 8:00  
CYCLE COMPLETE 8:25  
-----

Digital Signature #  
8079BFD2926690CF  
-----