

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000522  
12:40 28/05/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:07  
PRESSURIZING 3:11  
STERILIZING 4:21  
136.0 C 228kPa 4:21  
136.4 C 229kPa 4:51  
136.3 C 230kPa 5:21  
136.2 C 232kPa 5:51  
136.4 C 232kPa 6:21  
136.2 C 232kPa 6:51  
136.2 C 231kPa 7:21  
136.4 C 231kPa 7:51  
136.4 C 231kPa 7:51  
VENTING 7:51  
AIR DRYING 8:48  
CYCLE COMPLETE 9:24  
-----

Digital Signature #  
A308DBB785FB3F5D  
-----