

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2156

Name: RITTER 1  
Sterilizer ID: V2178198  
07/20/2022 12:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.8 F 0.0  
2:00 126.1 F 0.1  
4:00 149.7 F 2.9  
6:00 172.9 F 6.6  
8:00 198.7 F 11.0  
10:00 215.5 F 4.2  
12:00 225.2 F 8.9  
14:00 236.1 F 13.4  
16:00 236.5 F 9.9  
18:00 238.3 F 8.9  
20:00 246.2 F 13.6  
22:00 257.8 F 19.8  
24:00 268.3 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.6  
1:00 271.8 F 29.0  
2:00 272.4 F 29.2  
3:00 271.5 F 28.5  
4:00 272.2 F 29.3

Min 271.4 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.4 F 0.0  
5:00 229.9 F 0.0  
10:00 228.6 F 0.0  
15:00 230.8 F 0.0  
20:00 227.7 F 0.0  
25:00 235.6 F 0.0  
30:00 230.9 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 25:08  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:03:09  
CYCLE COMPLETE

07/20/2022 01:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_