
Midmark M11 - v1.0.5
Total Cycles: 1051

Name: RITTER2
Sterilizer ID: V2313228
05/02/2022 06:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.6 F 0.0
2:00 76.7 F 0.0
4:00 106.2 F 0.0
6:00 125.0 F 0.0
8:00 153.7 F 0.3
10:00 173.9 F 2.6
12:00 196.0 F 6.0
14:00 206.7 F 9.8
16:00 219.4 F 10.7
18:00 227.3 F 7.9
20:00 237.8 F 12.3
22:00 235.8 F 9.6
24:00 239.3 F 9.7
26:00 247.2 F 14.2
28:00 258.5 F 20.4
30:00 268.9 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.4 F 29.2
2:00 272.2 F 29.6
3:00 271.9 F 28.9
4:00 272.6 F 29.6

Min 271.7 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 199.4 F 0.0
10:00 208.6 F 0.0
15:00 217.0 F 0.0
20:00 220.6 F 0.0
25:00 217.5 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 31:09
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:27
CYCLE COMPLETE

05/02/2022 08:08 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____