

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4876

Name: RITTER 1  
Sterilizer ID: V2178198  
09/06/2024 04:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.7 F 0.0  
2:00 177.3 F 0.9  
4:00 205.2 F 6.2  
6:00 222.8 F 12.1  
8:00 228.0 F 7.7  
10:00 240.3 F 12.1  
12:00 242.3 F 11.9  
14:00 241.2 F 10.8  
16:00 256.0 F 18.5  
18:00 268.8 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 272.1 F 28.3  
2:00 272.6 F 28.7  
3:00 272.8 F 29.1  
4:00 271.8 F 28.3  
  
Min 271.7 F 28.1  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.8 F 0.0  
5:00 213.1 F 0.0  
10:00 222.7 F 0.0  
15:00 225.4 F 0.0  
20:00 223.6 F 0.0  
25:00 221.2 F 0.0  
30:00 220.2 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 18:60  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:35  
CYCLE COMPLETE

09/06/2024 05:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_