
Midmark M11 - v1.0.5
Total Cycles: 1976

Name: RITTER2
Sterilizer ID: V2313228
01/30/2023 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.6 F 0.0
2:00 165.7 F 0.5
4:00 170.2 F 4.1
6:00 195.9 F 8.8
8:00 218.3 F 9.7
10:00 226.6 F 9.6
12:00 233.8 F 8.9
14:00 240.7 F 12.4
16:00 240.9 F 11.1
18:00 253.6 F 17.8
20:00 265.0 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.7 F 30.0
2:00 272.0 F 29.3
3:00 271.7 F 29.6
4:00 272.7 F 29.9

Min 271.4 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 215.0 F 0.0
10:00 219.7 F 0.0
15:00 227.3 F 0.0
20:00 230.5 F 0.0
25:00 232.3 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:53
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:12
CYCLE COMPLETE

01/30/2023 10:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____