
Midmark M11 - v1.0.0
Total Cycles: 3043

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2023 08:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.3 F 0.0
2:00 164.5 F 1.9
4:00 171.8 F 5.5
6:00 195.0 F 9.7
8:00 218.4 F 6.2
10:00 225.1 F 9.0
12:00 237.7 F 10.4
14:00 239.3 F 11.1
16:00 236.7 F 9.0
18:00 250.8 F 15.7
20:00 262.5 F 22.3
22:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 28.8
2:00 271.3 F 28.4
3:00 272.5 F 28.9
4:00 271.3 F 28.2

Min 271.2 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 215.7 F 0.0
10:00 220.8 F 0.0
15:00 224.5 F 0.0
20:00 229.3 F 0.0
25:00 228.7 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:19
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:00:33
CYCLE COMPLETE

03/28/2023 09:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____