
Midmark M11 - v1.0.5
Total Cycles: 1601

Name: RITTER2
Sterilizer ID: V2313228
10/14/2022 03:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.6 F 0.0
2:00 124.6 F 0.9
4:00 143.6 F 3.5
6:00 175.1 F 7.6
8:00 197.4 F 12.6
10:00 214.5 F 5.7
12:00 227.1 F 10.1
14:00 232.0 F 8.6
16:00 241.1 F 12.3
18:00 240.1 F 10.9
20:00 252.9 F 17.3
22:00 264.2 F 24.1
24:00 271.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 273.0 F 29.8
2:00 271.9 F 29.0
3:00 272.1 F 29.6
4:00 272.6 F 29.6

Min 271.5 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 200.7 F 0.0
10:00 215.9 F 0.0
15:00 220.7 F 0.0
20:00 221.0 F 0.0
25:00 220.6 F 0.0
30:00 219.2 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:02
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:16
CYCLE COMPLETE

10/14/2022 04:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____