

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2317

Name: RITTER2  
Sterilizer ID: V2313228  
05/01/2023 10:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
2:00 191.2 F 3.0  
4:00 187.4 F 7.5  
6:00 208.4 F 13.1  
8:00 224.7 F 8.2  
10:00 235.0 F 11.8  
12:00 239.1 F 11.9  
14:00 239.1 F 10.9  
16:00 252.8 F 17.7  
18:00 264.4 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.4  
1:00 272.4 F 30.1  
2:00 271.7 F 29.4  
3:00 271.7 F 29.9  
4:00 272.3 F 30.0  
  
Min 271.1 F 29.2  
Max 272.5 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.1 F 0.0  
5:00 215.8 F 0.0  
10:00 222.7 F 0.0  
15:00 230.6 F 0.0  
20:00 229.1 F 0.0  
25:00 228.9 F 0.0  
30:00 221.6 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 19:59  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:58:55  
CYCLE COMPLETE

05/01/2023 11:16 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_