

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1311

Name: RITTER 1  
Sterilizer ID: V2178198  
10/28/2021 06:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.0 F 0.0  
2:00 85.3 F 0.0  
4:00 95.3 F 0.0  
6:00 113.6 F 0.0  
8:00 140.9 F 0.2  
10:00 167.5 F 2.2  
12:00 186.4 F 5.1  
14:00 200.8 F 8.2  
16:00 213.9 F 11.6  
18:00 221.1 F 5.4  
20:00 232.7 F 10.0  
22:00 236.4 F 9.3  
24:00 243.0 F 12.2  
26:00 240.3 F 10.5  
28:00 253.7 F 17.1  
30:00 264.7 F 23.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.6 F 28.7  
2:00 271.6 F 28.4  
3:00 272.4 F 28.6  
4:00 271.7 F 28.4  
  
Min 271.4 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.3 F 0.0  
5:00 187.6 F 0.0  
10:00 203.7 F 0.0  
15:00 203.7 F 0.0  
20:00 200.7 F 0.0  
25:00 196.3 F 0.0  
30:00 187.2 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 31:53  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:09:55  
CYCLE COMPLETE

10/28/2021 08:04 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02