

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 002525  
9:23 26/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:29  
PRESSURIZING 3:30  
STERILIZING 4:41  
136.4 C 220kPa 4:41  
136.3 C 223kPa 5:11  
136.4 C 223kPa 5:41  
136.4 C 224kPa 6:11  
136.4 C 224kPa 6:41  
136.2 C 221kPa 7:11  
136.4 C 220kPa 7:41  
136.3 C 222kPa 8:11  
136.3 C 222kPa 8:11  
VENTING 8:11  
AIR DRYING 9:07  
CYCLE COMPLETE 9:09

-----  
Digital Signature #  
6DB65CF10351B2DD  
-----