
Midmark M11 - v1.0.0
Total Cycles: 3834

Name: RITTER 1
Sterilizer ID: V2178198
09/28/2023 11:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.1 F 0.0
2:00 130.9 F 0.0
4:00 151.4 F 1.1
6:00 181.0 F 4.3
8:00 202.3 F 8.4
10:00 219.6 F 13.5
12:00 226.5 F 8.3
14:00 238.2 F 13.3
16:00 238.7 F 10.5
18:00 235.7 F 8.4
20:00 250.7 F 15.4
22:00 262.5 F 22.1
24:00 272.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.0 F 28.1
2:00 272.5 F 28.6
3:00 271.7 F 28.4
4:00 271.9 F 28.1

Min 271.4 F 27.8
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.1 F 0.0
5:00 206.8 F 0.0
10:00 213.8 F 0.0
15:00 219.5 F 0.0
20:00 227.9 F 0.0
25:00 231.3 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 24:13
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:02:21
CYCLE COMPLETE

09/28/2023 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____