
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005235
12:26 15/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:34
STERILIZING 3:28
136.5 C 224kPa 3:28
136.4 C 223kPa 3:58
136.4 C 223kPa 4:28
136.4 C 224kPa 4:58
136.3 C 222kPa 5:28
136.2 C 224kPa 5:58
136.3 C 222kPa 6:28
136.2 C 224kPa 6:58
136.2 C 224kPa 6:58
VENTING 6:58
AIR DRYING 7:48
CYCLE COMPLETE 10:10

Digital Signature #
C37A98042D169615
