
Midmark M11 - v1.0.5
Total Cycles: 3327

Name: RITTER2
Sterilizer ID: V2313228
01/26/2024 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.4 F 0.0
2:00 137.7 F 1.1
4:00 155.3 F 4.0
6:00 184.3 F 8.2
8:00 201.1 F 13.3
10:00 220.7 F 7.2
12:00 233.0 F 12.2
14:00 234.0 F 9.1
16:00 239.2 F 10.0
18:00 246.5 F 13.9
20:00 258.7 F 20.6
22:00 269.7 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.5
2:00 271.4 F 29.0
3:00 272.6 F 29.7
4:00 271.7 F 29.0

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 225.2 F 0.0
10:00 202.1 F 0.0
15:00 200.5 F 0.0
20:00 199.7 F 0.0
25:00 200.8 F 0.0
30:00 205.9 F 0.0
Drying Complete

FILLING: 0:38
HEATING: 22:59
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:53
CYCLE COMPLETE

01/26/2024 03:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____