
Midmark M11 - v1.0.0
Total Cycles: 1962

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2022 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.5 F 0.0
2:00 156.8 F 2.3
4:00 162.3 F 5.7
6:00 195.4 F 9.9
8:00 214.1 F 5.3
10:00 226.2 F 7.8
12:00 235.7 F 12.6
14:00 234.1 F 9.1
16:00 240.3 F 9.9
18:00 245.1 F 13.1
20:00 257.2 F 19.5
22:00 268.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.3
2:00 271.3 F 28.5
3:00 272.6 F 29.4
4:00 272.0 F 28.8

Min 271.3 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 222.6 F 0.0
10:00 225.5 F 0.0
15:00 226.6 F 0.0
20:00 228.7 F 0.0
25:00 231.4 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:19
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:27
CYCLE COMPLETE

05/27/2022 09:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____