

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3032

Name: RITTER2  
Sterilizer ID: V2313228  
10/19/2023 12:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.7 F 0.0  
2:00 178.5 F 3.1  
4:00 184.5 F 7.4  
6:00 206.0 F 12.3  
8:00 220.0 F 7.1  
10:00 233.1 F 12.2  
12:00 234.9 F 9.7  
14:00 237.4 F 9.3  
16:00 248.7 F 15.3  
18:00 260.8 F 22.1  
20:00 271.8 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 272.8 F 29.9  
2:00 271.9 F 29.3  
3:00 271.5 F 29.5  
4:00 272.5 F 29.9  
  
Min 271.2 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
5:00 216.2 F 0.0  
10:00 217.5 F 0.0  
15:00 218.1 F 0.0  
20:00 216.4 F 0.0  
25:00 215.5 F 0.0  
30:00 214.2 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 20:37  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 00:59:45  
CYCLE COMPLETE

10/19/2023 01:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_