
Midmark M11 - v1.0.5
Total Cycles: 3604

Name: RITTER2
Sterilizer ID: V2313228
05/10/2024 08:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 81.5 F 0.0
4:00 100.6 F 0.0
6:00 112.5 F 1.5
8:00 146.2 F 4.6
10:00 172.1 F 7.8
12:00 201.9 F 11.6
14:00 218.6 F 5.0
16:00 224.9 F 9.0
18:00 234.6 F 13.2
20:00 234.8 F 9.2
22:00 239.6 F 10.7
24:00 243.7 F 12.6
26:00 255.7 F 18.7
28:00 266.0 F 25.1
30:00 272.0 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.5
1:00 272.0 F 29.4
2:00 272.2 F 29.3
3:00 272.6 F 29.5
4:00 273.0 F 29.7

Min 271.7 F 29.3
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 227.1 F 0.0
10:00 186.5 F 0.0
15:00 178.2 F 0.0
20:00 176.1 F 0.0
25:00 173.9 F 0.0
30:00 172.4 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:27
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:34
CYCLE COMPLETE

05/10/2024 09:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____