
Midmark M11 - v1.0.5
Total Cycles: 3164

Name: RITTER2
Sterilizer ID: V2313228
11/29/2023 08:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.2 F 0.0
2:00 82.0 F 0.0
4:00 74.0 F 0.2
6:00 102.0 F 1.4
8:00 134.2 F 3.8
10:00 165.3 F 6.9
12:00 188.0 F 10.4
14:00 214.5 F 4.6
16:00 214.6 F 6.7
18:00 231.0 F 10.7
20:00 229.6 F 7.7
22:00 240.1 F 12.0
24:00 237.3 F 9.9
26:00 250.7 F 16.2
28:00 262.2 F 22.7
30:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.1 F 29.4
2:00 271.8 F 29.5
3:00 271.9 F 29.1
4:00 272.0 F 29.8

Min 271.2 F 28.9
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.5 F 0.0
5:00 191.6 F 0.0
10:00 183.8 F 0.0
15:00 181.6 F 0.0
20:00 178.6 F 0.0
25:00 175.9 F 0.0
30:00 173.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:23
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:23
CYCLE COMPLETE

11/29/2023 09:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____