
Midmark M11 - v1.0.0
Total Cycles: 2621

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2022 03:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.5 F 0.0
2:00 153.4 F 3.2
4:00 176.6 F 7.2
6:00 199.3 F 12.0
8:00 214.8 F 5.5
10:00 226.3 F 10.7
12:00 230.0 F 6.6
14:00 244.5 F 12.8
16:00 241.8 F 11.1
18:00 254.8 F 17.6
20:00 266.0 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.4 F 28.8
2:00 271.2 F 28.0
3:00 272.3 F 28.9
4:00 272.1 F 28.4

Min 271.1 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 236.7 F 0.0
10:00 231.5 F 0.0
15:00 228.4 F 0.0
20:00 232.3 F 0.0
25:00 229.7 F 0.0
30:00 231.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 21:40
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

11/29/2022 04:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____