
Midmark M11 - v1.0.0
Total Cycles: 3586

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2023 11:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.0 F 0.0
2:00 153.3 F 0.4
4:00 171.8 F 2.6
6:00 193.1 F 5.8
8:00 208.0 F 9.9
10:00 218.4 F 8.2
12:00 227.8 F 8.6
14:00 239.0 F 13.3
16:00 238.8 F 10.5
18:00 236.0 F 8.4
20:00 249.6 F 14.9
22:00 260.8 F 21.2
24:00 271.0 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.9
2:00 271.7 F 28.2
3:00 271.9 F 28.7
4:00 272.3 F 28.7

Min 271.4 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 218.8 F 0.0
10:00 214.1 F 0.0
15:00 196.2 F 0.0
20:00 194.9 F 0.0
25:00 202.5 F 0.0
30:00 209.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 24:46
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:03:06
CYCLE COMPLETE

07/26/2023 12:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____