
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000896
7:04 30/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:34
STERILIZING 4:39
136.0 C 220kPa 4:39
136.1 C 221kPa 5:09
136.2 C 222kPa 5:39
136.3 C 222kPa 6:09
136.3 C 222kPa 6:39
136.3 C 222kPa 7:09
136.4 C 222kPa 7:39
135.9 C 219kPa 8:09
135.9 C 219kPa 8:09
VENTING 8:09
AIR DRYING 9:02
CYCLE COMPLETE 9:03

Digital Signature #
5FC788CAB945B4B0
