
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001388
14:00 16/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:34
STERILIZING 3:42
136.5 C 223kPa 3:42
136.2 C 224kPa 4:12
136.4 C 226kPa 4:42
136.0 C 222kPa 5:12
136.2 C 224kPa 5:42
136.3 C 225kPa 6:12
136.4 C 226kPa 6:42
136.3 C 224kPa 7:12
136.3 C 224kPa 7:12
VENTING 7:12
AIR DRYING 8:04
CYCLE COMPLETE 23:05

Digital Signature #
7CE8FD4547F49DB1
