
Midmark M11 - v1.0.5
Total Cycles: 3154

Name: RITTER2
Sterilizer ID: V2313228
11/22/2023 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.3 F 0.0
2:00 103.6 F 0.2
4:00 127.9 F 1.7
6:00 153.6 F 4.5
8:00 183.1 F 8.0
10:00 198.4 F 12.3
12:00 212.3 F 4.9
14:00 226.7 F 8.8
16:00 238.4 F 11.5
18:00 238.2 F 11.3
20:00 237.8 F 10.0
22:00 250.5 F 16.2
24:00 261.8 F 22.8
26:00 272.4 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.9 F 29.8
2:00 271.8 F 29.3
3:00 273.0 F 29.8
4:00 271.7 F 29.2

Min 271.7 F 29.1
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 195.9 F 0.0
10:00 202.9 F 0.0
15:00 202.3 F 0.0
20:00 200.7 F 0.0
25:00 199.4 F 0.0
30:00 199.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 26:28
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:04:46
CYCLE COMPLETE

11/22/2023 01:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____