
Midmark M11 - v1.0.5
Total Cycles: 3807

Name: RITTER2
Sterilizer ID: V2313228
07/23/2024 09:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 69.7 F 0.0
4:00 75.8 F 0.0
6:00 100.0 F 0.1
8:00 141.5 F 1.0
10:00 166.5 F 3.5
12:00 186.9 F 6.0
14:00 199.6 F 9.3
16:00 208.2 F 12.7
18:00 218.1 F 6.3
20:00 228.4 F 10.2
22:00 233.1 F 9.3
24:00 238.5 F 11.4
26:00 236.4 F 9.3
28:00 248.6 F 15.2
30:00 259.0 F 20.9
32:00 268.7 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 29.9
2:00 272.0 F 29.2
3:00 271.4 F 29.4
4:00 272.7 F 29.8

Min 271.2 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 188.3 F 0.0
10:00 192.5 F 0.0
15:00 183.4 F 0.0
20:00 184.0 F 0.0
25:00 182.6 F 0.0
30:00 182.9 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 33:18
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:11:50
CYCLE COMPLETE

07/23/2024 10:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____