
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002173
11:17 28/10/2024
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 4:36
STERILIZING 5:23
136.4 C 226kPa 5:23
136.4 C 222kPa 5:53
136.4 C 223kPa 6:23
135.8 C 221kPa 6:53
136.1 C 222kPa 7:23
136.4 C 223kPa 7:53
136.4 C 223kPa 8:23
136.0 C 222kPa 8:53
136.4 C 224kPa 9:23
136.4 C 224kPa 9:53
136.3 C 225kPa 10:23
136.4 C 225kPa 10:53
136.4 C 225kPa 11:23
136.2 C 223kPa 11:53
136.2 C 221kPa 12:23
136.2 C 224kPa 12:53
136.5 C 225kPa 13:23
136.4 C 225kPa 13:53
136.1 C 223kPa 14:23
136.4 C 226kPa 14:53
136.4 C 225kPa 15:23
136.4 C 225kPa 15:23
VENTING 15:23
AIR DRYING 16:20
CYCLE COMPLETE 16:28

Digital Signature #
E09D0B12815AF943
