
Midmark M11 - v1.0.0
Total Cycles: 1336

Name: RITTER 1
Sterilizer ID: V2178198
11/03/2021 01:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 90.7 F 0.0
2:00 92.0 F 0.0
4:00 102.5 F 0.1
6:00 134.2 F 0.1
8:00 188.0 F 0.2
10:00 206.3 F 2.8
12:00 216.3 F 5.8
14:00 226.7 F 8.9
16:00 237.5 F 12.7
18:00 234.4 F 8.6
20:00 244.8 F 12.8
22:00 242.7 F 11.4
24:00 239.5 F 9.8
26:00 252.8 F 16.3
28:00 263.2 F 22.3
30:00 272.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.8
1:00 272.6 F 28.6
2:00 272.0 F 28.0
3:00 271.4 F 28.0
4:00 272.7 F 28.7

Min 271.2 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 225.2 F 0.0
10:00 224.2 F 0.0
15:00 225.7 F 0.0
20:00 216.3 F 0.0
25:00 212.8 F 0.0
30:00 212.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 30:23
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:08:39
CYCLE COMPLETE

11/03/2021 03:05 PM
ACCEPTED: X
REJECTED:
Operator: U01