
Midmark M11 - v1.0.0
Total Cycles: 3170

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2023 02:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.1 F 0.0
2:00 130.8 F 0.0
4:00 162.5 F 1.4
6:00 188.4 F 5.1
8:00 207.9 F 9.8
10:00 221.4 F 7.6
12:00 230.1 F 9.5
14:00 237.3 F 9.4
16:00 242.7 F 12.1
18:00 240.5 F 10.5
20:00 254.0 F 17.2
22:00 265.7 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.2 F 28.6
2:00 272.1 F 28.8
3:00 271.8 F 28.3
4:00 272.6 F 29.0

Min 271.3 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 204.5 F 0.0
10:00 216.1 F 0.0
15:00 230.7 F 0.0
20:00 223.9 F 0.0
25:00 220.9 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:51
CYCLE COMPLETE

04/27/2023 03:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____