

-----  
Midmark M11 - v1.0.5  
Total Cycles: 310

Name: RITTER2  
Sterilizer ID: V2313228  
08/25/2021 09:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.5 F 0.0  
2:00 183.4 F 3.7  
4:00 201.4 F 9.7  
6:00 220.6 F 6.5  
8:00 235.6 F 12.0  
10:00 237.3 F 10.4  
12:00 237.6 F 9.8  
14:00 252.5 F 17.2  
16:00 264.9 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 272.1 F 29.1  
2:00 272.7 F 29.5  
3:00 272.7 F 29.9  
4:00 271.9 F 29.0  
  
Min 271.8 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.2 F 0.0  
5:00 196.2 F 0.0  
10:00 193.6 F 0.0  
15:00 188.2 F 0.0  
20:00 185.1 F 0.0  
25:00 182.8 F 0.0  
30:00 180.3 F 0.0  
  
Drying Complete

FILLING: 2:43  
HEATING: 17:42  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:57:30  
CYCLE COMPLETE

08/25/2021 10:39 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02