
Midmark M11 - v1.0.0
Total Cycles: 5481

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2025 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.8 F 0.0
2:00 167.4 F 3.9
4:00 183.0 F 8.9
6:00 208.3 F 7.1
8:00 224.2 F 8.3
10:00 237.2 F 11.6
12:00 238.7 F 10.4
14:00 234.5 F 7.8
16:00 249.9 F 15.0
18:00 262.3 F 22.1
20:00 272.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 271.6 F 28.1
2:00 271.9 F 28.1
3:00 272.3 F 28.3
4:00 272.1 F 28.6

Min 271.4 F 28.0
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 211.4 F 0.0
10:00 200.8 F 0.0
15:00 202.5 F 0.0
20:00 208.3 F 0.0
25:00 238.7 F 0.0
30:00 243.6 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

04/04/2025 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____