
Midmark M11 - v1.0.5
Total Cycles: 697

Name: RITTER2
Sterilizer ID: V2313228
01/13/2022 02:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.8 F 0.0
2:00 171.1 F 3.3
4:00 194.3 F 7.8
6:00 215.2 F 13.5
8:00 229.3 F 8.8
10:00 236.1 F 10.5
12:00 242.7 F 12.7
14:00 242.7 F 12.0
16:00 256.1 F 19.0
18:00 267.9 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.3
1:00 272.2 F 29.2
2:00 272.6 F 29.3
3:00 272.4 F 29.6
4:00 272.5 F 29.9

Min 271.8 F 29.2
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.7 F 0.0
5:00 226.5 F 0.0
10:00 198.8 F 0.0
15:00 192.7 F 0.0
20:00 188.6 F 0.0
25:00 185.5 F 0.0
30:00 184.1 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 19:56
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

01/13/2022 03:31 PM
ACCEPTED: X
REJECTED:
Operator: U02