
Midmark M11 - v1.0.0
Total Cycles: 5195

Name: RITTER 1
Sterilizer ID: V2178198
12/16/2024 04:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 218.2 F 0.0
2:00 210.0 F 4.6
4:00 224.1 F 11.0
6:00 227.3 F 6.8
8:00 243.1 F 13.1
10:00 241.4 F 11.0
12:00 239.0 F 9.6
14:00 254.7 F 17.4
16:00 267.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 27.9
1:00 272.6 F 28.4
2:00 272.5 F 28.7
3:00 271.6 F 27.8
4:00 272.6 F 28.4

Min 271.5 F 27.7
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 210.2 F 0.0
10:00 214.7 F 0.0
15:00 213.3 F 0.0
20:00 210.7 F 0.0
25:00 217.6 F 0.0
30:00 211.6 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 17:05
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:55:56
CYCLE COMPLETE

12/16/2024 05:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____