
Midmark M11 - v1.0.0
Total Cycles: 2144

Name: RITTER 1
Sterilizer ID: V2178198
07/18/2022 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.4 F 0.0
2:00 165.2 F 1.4
4:00 179.0 F 5.3
6:00 207.7 F 10.1
8:00 223.9 F 5.7
10:00 234.3 F 10.7
12:00 233.7 F 7.1
14:00 244.2 F 13.5
16:00 242.9 F 12.0
18:00 255.8 F 18.6
20:00 266.9 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.6 F 29.5
2:00 272.1 F 28.8
3:00 271.4 F 28.8
4:00 272.5 F 29.5

Min 271.2 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 228.3 F 0.0
10:00 230.3 F 0.0
15:00 230.6 F 0.0
20:00 233.7 F 0.0
25:00 221.1 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 21:33
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

07/18/2022 12:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____