
Midmark M11 - v1.0.5
Total Cycles: 2188

Name: RITTER2
Sterilizer ID: V2313228
03/28/2023 03:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.1 F 0.0
2:00 176.8 F 3.9
4:00 201.9 F 8.5
6:00 217.4 F 9.6
8:00 229.1 F 8.8
10:00 235.2 F 10.1
12:00 242.2 F 12.4
14:00 242.0 F 11.6
16:00 255.8 F 18.8
18:00 267.7 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.8 F 29.9
2:00 272.1 F 29.2
3:00 271.5 F 29.3
4:00 272.6 F 29.9

Min 271.3 F 28.9
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.1 F 0.0
5:00 233.7 F 0.0
10:00 227.0 F 0.0
15:00 225.3 F 0.0
20:00 231.6 F 0.0
25:00 223.9 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

03/28/2023 04:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____