
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001499
9:09 05/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:13
STERILIZING 4:21
136.4 C 223kPa 4:21
136.4 C 223kPa 4:51
136.5 C 224kPa 5:21
136.2 C 220kPa 5:51
136.4 C 220kPa 6:21
136.0 C 220kPa 6:51
136.0 C 221kPa 7:21
136.4 C 221kPa 7:51
136.4 C 221kPa 7:51
VENTING 7:51
AIR DRYING 8:41
CYCLE COMPLETE 10:25

Digital Signature #
6FEB68C160EABF47
