
Midmark M11 - v1.0.5
Total Cycles: 1842

Name: RITTER2
Sterilizer ID: V2313228
12/14/2022 03:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.9 F 0.0
2:00 197.5 F 4.6
4:00 212.6 F 10.7
6:00 224.5 F 6.8
8:00 239.2 F 13.0
10:00 240.6 F 11.6
12:00 241.0 F 11.1
14:00 255.3 F 18.5
16:00 267.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.8
1:00 272.7 F 29.5
2:00 271.5 F 28.8
3:00 272.5 F 29.4
4:00 271.6 F 28.7

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 226.3 F 0.0
10:00 232.2 F 0.0
15:00 227.0 F 0.0
20:00 223.4 F 0.0
25:00 223.8 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 2:37
HEATING: 17:09
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:56:57
CYCLE COMPLETE

12/14/2022 04:43 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____