

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4401

Name: RITTER 1  
Sterilizer ID: V2178198  
03/25/2024 10:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.2 F 0.0  
2:00 164.6 F 1.5  
4:00 158.4 F 4.3  
6:00 174.0 F 8.3  
8:00 199.3 F 12.3  
10:00 219.5 F 5.5  
12:00 225.0 F 10.2  
14:00 230.4 F 7.7  
16:00 238.5 F 11.8  
18:00 238.9 F 9.8  
20:00 251.6 F 16.0  
22:00 262.4 F 22.4  
24:00 272.5 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.5 F 29.0  
1:00 272.7 F 29.4  
2:00 272.1 F 29.4  
3:00 272.4 F 29.0  
4:00 272.7 F 29.5

Min 271.8 F 28.9  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.1 F 0.0  
5:00 233.3 F 0.0  
10:00 231.5 F 0.0  
15:00 228.4 F 0.0  
20:00 227.1 F 0.0  
25:00 225.7 F 0.0  
30:00 224.5 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 25:04  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:03:37  
CYCLE COMPLETE

03/25/2024 11:13 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_