
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001613
12:56 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:36
STERILIZING 3:41
136.5 C 225kPa 3:41
136.4 C 226kPa 4:11
136.4 C 226kPa 4:41
136.4 C 225kPa 5:11
136.4 C 227kPa 5:41
136.4 C 223kPa 6:11
136.3 C 225kPa 6:41
136.4 C 226kPa 7:11
136.4 C 226kPa 7:11
VENTING 7:11
AIR DRYING 8:12
CYCLE COMPLETE 8:19

Digital Signature #
9ECA3CA7DF0BD16D
