
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001227
16:12 23/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:25
STERILIZING 3:28
136.0 C 220kPa 3:28
136.4 C 225kPa 3:58
136.2 C 222kPa 4:28
136.3 C 223kPa 4:58
136.3 C 222kPa 5:28
136.4 C 219kPa 5:58
136.4 C 223kPa 6:28
136.3 C 222kPa 6:58
136.3 C 222kPa 6:58
VENTING 6:59
AIR DRYING 7:45
CYCLE COMPLETE 11:47

Digital Signature #
82DE7BD7D522F175
