
Midmark M11 - v1.0.5
Total Cycles: 4334

Name: RITTER2
Sterilizer ID: V2313228
02/05/2025 11:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.5 F 0.0
2:00 136.1 F 1.2
4:00 136.8 F 3.4
6:00 173.9 F 6.9
8:00 188.9 F 11.2
10:00 212.9 F 4.4
12:00 223.9 F 8.8
14:00 235.6 F 13.5
16:00 237.5 F 10.4
18:00 236.5 F 8.6
20:00 248.6 F 14.8
22:00 259.6 F 21.2
24:00 270.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 273.1 F 29.7
1:00 272.5 F 30.1
2:00 272.4 F 29.7
3:00 272.6 F 30.0
4:00 272.5 F 29.6

Min 272.2 F 29.6
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 206.5 F 0.0
10:00 204.5 F 0.0
15:00 202.3 F 0.0
20:00 213.6 F 0.0
25:00 217.6 F 0.0
30:00 197.4 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 25:18
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:03:52
CYCLE COMPLETE

02/05/2025 12:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____