
Midmark M11 - v1.0.0
Total Cycles: 2483

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2022 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.9 F 0.0
2:00 121.9 F 0.7
4:00 154.2 F 3.3
6:00 175.3 F 6.6
8:00 196.0 F 10.5
10:00 219.1 F 5.3
12:00 226.3 F 7.2
14:00 230.0 F 11.7
16:00 231.1 F 7.6
18:00 244.2 F 13.0
20:00 240.0 F 10.8
22:00 253.1 F 16.9
24:00 263.8 F 23.3
26:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.0
2:00 272.4 F 29.4
3:00 271.2 F 28.7
4:00 272.7 F 29.2

Min 271.2 F 28.6
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 216.7 F 0.0
10:00 223.1 F 0.0
15:00 228.0 F 0.0
20:00 201.0 F 0.0
25:00 183.8 F 0.0
30:00 170.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:53
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:05:04
CYCLE COMPLETE

10/25/2022 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____