
Midmark M11 - v1.0.5
Total Cycles: 789

Name: RITTER2
Sterilizer ID: V2313228
02/11/2022 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.6 F 0.0
2:00 123.3 F 0.8
4:00 133.3 F 3.3
6:00 160.7 F 7.2
8:00 187.1 F 12.3
10:00 213.3 F 5.5
12:00 227.7 F 10.8
14:00 231.3 F 7.9
16:00 242.5 F 11.9
18:00 244.5 F 13.0
20:00 256.8 F 19.6
22:00 268.2 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 271.9 F 29.2
2:00 272.1 F 29.4
3:00 272.3 F 29.6
4:00 272.3 F 29.6

Min 271.7 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 233.9 F 0.0
5:00 204.4 F 0.0
10:00 202.9 F 0.0
15:00 202.9 F 0.0
20:00 200.2 F 0.0
25:00 197.5 F 0.0
30:00 195.6 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 23:07
STERILIZING: 4:00
VENTING: 2:44
DRYING: 30:00
TOTAL CYCLE: 01:00:58
CYCLE COMPLETE

02/11/2022 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____