
Midmark M11 - v1.0.0
Total Cycles: 3952

Name: RITTER 1
Sterilizer ID: V2178198
10/27/2023 01:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.6 F 0.0
2:00 154.7 F 2.6
4:00 183.4 F 6.4
6:00 203.4 F 10.5
8:00 211.7 F 3.8
10:00 223.3 F 8.6
12:00 235.3 F 13.3
14:00 236.4 F 10.0
16:00 236.7 F 8.6
18:00 247.6 F 14.2
20:00 259.4 F 20.6
22:00 270.1 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.5
1:00 272.3 F 29.0
2:00 271.7 F 28.8
3:00 271.9 F 28.6
4:00 272.2 F 29.2

Min 271.4 F 28.3
Max 272.3 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 218.9 F 0.0
10:00 223.3 F 0.0
15:00 226.1 F 0.0
20:00 225.1 F 0.0
25:00 226.0 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 22:41
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:18
CYCLE COMPLETE

10/27/2023 02:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____