
Midmark M11 - v1.0.0
Total Cycles: 5146

Name: RITTER 1
Sterilizer ID: V2178198
12/02/2024 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.7 F 0.0
2:00 86.3 F 0.0
4:00 91.0 F 0.0
6:00 112.8 F 0.0
8:00 144.1 F 0.2
10:00 171.3 F 0.3
12:00 193.2 F 0.6
14:00 209.2 F 2.5
16:00 223.1 F 5.7
18:00 231.4 F 8.8
20:00 238.7 F 12.5
22:00 235.8 F 8.6
24:00 244.7 F 12.2
26:00 243.9 F 11.9
28:00 240.4 F 10.3
30:00 253.3 F 16.6
32:00 263.9 F 22.9
34:00 272.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.4 F 28.5
2:00 272.7 F 28.6
3:00 272.0 F 28.0
4:00 271.7 F 28.0

Min 271.6 F 27.8
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 218.8 F 0.0
10:00 241.2 F 0.0
15:00 236.4 F 0.0
20:00 228.3 F 0.0
25:00 228.6 F 0.0
30:00 231.3 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 34:17
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:12:26
CYCLE COMPLETE

12/02/2024 10:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____